Name	Approved	Date
PA Eddy	PAE	2-15-83

Project Number _____

Internal Distribution

PE Bramson
PA Eddy
LS Prater
FILE
LB

Date February 18, 1983

To WJ Liechty

From JT Rieger JTR

Subject Technical Review of Offers - Purchase Requisition F6216

I have reviewed the three offers which you sent to me regarding the Liquid Level Monitoring System. None of these offers is technically acceptable as is. The following issues need to be addressed before I can determine whether they meet our requirements.

Branom Instrument Company - It is not clear whether the recorder which is provided will record in feet instead of meters.

Technical Controls, Inc. - Literature included with this offer says the system runs on 110V AC power, with no mention of conversion to the 12V DC power which we require. Also, there is no indication that a strip chart recorder is included in this system.

McCrometer-Ametek - This offer also neglects the conversion to 12V DC power.

In addition, we require that any of these systems be able to run for at least a week in the field without a battery charge.

As soon as I have information on these items, I can evaluate each system and let you know which ones are acceptable.

JTR:gd

54-1900-001 (3-71)
U S GOVERNMENT PRINTING OFFICE 1981—796-298/9

0I5BH000 0